

A11101 294945

NAT'L INST OF STANDARDS & TECH R.I.C.




A11101294945

/Bulletin of the Bureau of Standards
QC1 .U5 V7:1911 C.2 NBS-PUB-C 1905

DEPARTMENT OF COMMERCE AND LABOR
—
BULLETIN
OF THE
BUREAU OF STANDARDS
—
VOLUME 7
1911











DEPARTMENT OF COMMERCE AND LABOR

BULLETIN
OF THE
BUREAU OF STANDARDS

S. W. STRATTON, DIRECTOR

VOLUME 7
1911



WASHINGTON
GOVERNMENT PRINTING OFFICE
1911

A PHOTOMETRIC ATTACHMENT FOR SPECTROSCOPES

By P. G. Nutting

The instrument here described was designed for spectrophotometric work and with three ends in view: (1) High sensibility, (2) economy of light, and (3) variable dispersion. Tests of a sample instrument constructed by Fuess show that the sensibility has been pushed to the limit of sensibility (1.5 per cent) of




Fig. 1

the eye, while the light economy is quite satisfactory and the dispersion that of any spectroscope to which it is attached. No trace of stray light can be detected in the new instrument and the single face on which dirt could cause stray light is entirely inclosed. There is a slight correction to the scale reading, but this is less than 2 per cent at its maximum.

The attachment consists of two reflecting prisms, P_1 and P_2 , two Glan-Thompson nicols N_1 and N_2 and a lens arranged as shown in the figure. The whole is attached to the slit of any spectroscope, for example one of the Hilger "wave length" type. The essential feature is that a real image of the photometric surface (the common surface of the two reflecting prisms) is thrown on the slit by an achromatic lens and may thus be brought into the plane of the slit. The two beams of light to be compared, one passing through and the other reflected from the photometric surface, are brought to equality by rotating the nicol N_1 . The photometric surface is alternately silvered and clear, the silver strips being 0.25 mm wide and separated by clear strips equal in width to the silvered strips.

The combination of reflecting prisms was suggested by and is similar to that used by Dr. H. E. Ives¹ in his new spectrophotometer, but in his instrument the photometric surface is silvered halfway across instead of in fine strips.

Both beams are polarized by reflection and the instrument was carefully tested for departures from the cosine square law of intensities. First the reflecting system was removed and the elliptical polarization (change of phase and rotation) produced by reflection on each of the reflecting surfaces determined for 45° incidence and 45° azimuth. The second test was by means of an auxiliary nicol placed before the free opening so that both beams were polarized before reflection. The third test was with rotating sectorized disks of fixed known openings mounted before the free opening of the instrument. In this last test the sources were seasoned glow lamps operated in series with a rheostat on a storage battery circuit. The distances of these lamps were adjusted to equality of illumination with the 180° disk moving and the nicol at 0° . Then other sectors were substituted for the 180° sector and the readings of the nicol compared with their theoretical values deduced from the sector openings. The corrections were nowhere in excess of 2 per cent and of such sign as to indicate that they were in large part due to accidental errors.

The sensibility of the instrument was tested by rotating a parallel glass plate of known index in the path of one beam and

¹ Phys. Rev., 30, p. 450; 1910.

by repeating readings. Sensibilities varying from 0.015 to 0.018 were recorded.

The excellent showing made by the instrument is due in some measure to the excellent workmanship displayed in the construction of its optical parts. But two minor improvements suggest themselves; that the position of the lens be made adjustable from without and that the balancing nicol N_1 be provided with a larger (10 cm) divided circle to permit of more precise readings.

WASHINGTON, October 1, 1910.

